

Machine Learning Approaches for Automated Lesion Detection in Microwave Breast Imaging Clinical Data

Soumya Prakash Rana^{1,*}, Maitreyee Dey¹, Gianluigi Tiberi^{1,2}, Lorenzo Sani², Alessandro Vispa², Giovanni Raspa², Michele Duranti³, Mohammad Ghavami¹, and Sandra Dudley¹

¹Division of Electrical and Electronic Engineering, School of Engineering, London South Bank University, London, United Kingdom

²UBT Srl, Spin Off of the University of Perugia, Italy

³Department of Diagnostic Imaging, Perugia Hospital, Italy

*ranas9@lsbu.ac.uk

ABSTRACT

Breast lesion detection employing state of the art microwave systems provide a safe, non-ionizing technique that can differentiate healthy and non-healthy tissues by exploiting their dielectric properties. In this paper, a microwave apparatus for breast lesion detection is used to accumulate clinical data from subjects undergoing breast examinations at the Department of Diagnostic Imaging, Perugia Hospital, Perugia, Italy. This paper presents the first ever clinical demonstration and comparison of a microwave ultra-wideband (UWB) device augmented by machine learning with subjects who are simultaneously undergoing conventional breast examinations. Non-ionizing microwave signals are transmitted through the breast tissue and the scattering parameters (S-parameter) are received via a dedicated moving transmitting and receiving antenna set-up. The output of a parallel radiologist study for the same subjects, performed using conventional techniques, is taken to pre-process microwave data and create suitable data for the machine intelligence system. These data are used to train and investigate several suitable supervised machine learning algorithms nearest neighbour (NN), multi-layer perceptron (MLP) neural network, and support vector machine (SVM) to create an intelligent classification system towards supporting clinicians to recognise breasts with lesions. The results are rigorously analysed, validated through statistical measurements, and found the quadratic kernel of SVM can classify the breast data with 98% accuracy.

Introduction

Breast cancer is the most common cancer to affect women worldwide and the second most common cancer overall¹, with nearly 1.7 million new cases diagnosed annually². Mammography, the gold screening standard, is not suggested for screening women under 50 years old due to ionizing radiation exposure concerns. This means that 40% of all women in the EU (age 25-49 years old), representing 20% of breast cancer cases in Europe, cannot avail of the the most conventional breast cancer screening modality³. Furthermore, X-ray mammography cannot be undergone frequently, i.e. no more than once every 2 years in the EU, and it is prohibited for obvious reasons during pregnancy⁴. Breast cancer risk increases with further exposure to ionizing radiation from repeated mammography examinations⁵. Women who have undergone such tests also state that the exam is painful, particularly during their premenstrual period, or when the test is performed on women with smaller breasts⁶. Lastly, conventional mammography has been shown to miss approximately 15% of cancer (false negative)^{7,8}. Bearing in mind these limitations, new imaging approaches must be considered. Hitherto, microwave imaging has gained increased attention for its potential in breast cancer detection scenarios, fortified by the measurable variations in the dielectric properties of malignant and normal tissues at the microwave frequency ranges used. Explicitly, the work presented by Li, Xu, et al.⁹ and Bond, Essex J., et al.¹⁰, demonstrated that a substantial contrast between malignant and healthy breast tissue is present; this contrast was demonstrated to be up to a factor of five in conductivity and permittivity. More recent works propose that this contrast is only between malignant and fatty breast tissues, with a lower contrast (lower than 10% in dielectric properties) is found between healthy fibro glandular and malignant tissues¹¹⁻¹³. Moreover, Lazebnik, Mariya, et al., demonstrated that the dielectric properties of benign lesions are similar to the properties of fibro glandular tissues by¹³. Current microwave breast imaging research can be considered in two categories; microwave tomography and ultra-wideband (UWB) radar techniques¹⁴. A small number of prototypes are at clinical trial stages, including developments by Dartmouth College¹⁵ and the University of Bristol¹⁶. Specifically,¹⁵ employs microwave tomography and employs antennas with a matching liquid, while¹⁶ employs an UWB radar approach and uses an array of 60 antennas with a matching liquid.

34 A UWB microwave prototype (Mammowave UBT etc) has been constructed, tested and validated previously¹⁷. The system
35 operates in air employing two antennas and displays maps of dielectric property changes in tissues. Artefact removal (a
36 matching liquid is not used here) is performed through appropriate mathematical procedures^{18,19}. A Huygens Principle (HP)
37 approach is used to capture differences in dielectric properties and discriminates tissues and tissue condition. Test on phantoms
38 have shown a resolution of 1 mm¹⁸, while a sensitivity of 90% has been achieved in the ongoing clinical trial¹⁹.

39 Recently, machine learning based approaches for breast lesion detection have enjoyed increased attention^{20–22}. Machine
40 learning (ML) can be explicitly used to make decisions based on learned patterns (available datasets) and can automatically
41 create an analytical model for future predictions without direct human intervention. Various methods such as nearest neighbor,
42 neural networks, naive bayes, decision trees, conventional ML algorithms, and some hybrid approaches have been used for
43 classification purpose. Also, deep learning (DL) based methods for tumor classification has been investigated. However, limited
44 to breast lesion microwave imaging, ML and DL for breast lesion detection have been applied hitherto only to microwave
45 datasets obtained through numerical simulations or measurements in phantoms^{23–27} and nor ever before to clinical data.

46 The authors present the first ever work on clinically trialed UWB data augmented by ML for automated safe breast lesion
47 detection. The clinical trial UWB data have been collected at Perugia Hospital, Italy, using the microwave apparatus named
48 “MammoWave”, a non-ionizing and X-ray free mammogram invented by UBT Srl, Italy. In this research, we have investigated
49 the prospect of employing ML algorithms for computer-aided breast lesion detection to support clinicians, by reducing overhead
50 and increasing the speed in decision making between healthy and non-healthy lesion patterns from the clinically collected
51 data through the current microwave apparatus. Various ML algorithms have been applied in the field of pattern identification
52 and future prediction. Among them, three popular methods, k -nearest neighbor (k NN), multi-layer perceptron (MLP) neural
53 network, and support vector machine (SVM) are explored here to analyze the acquired labeled MammoWave data thoroughly.
54 These experiments have been performed to fit the labeled training data with the optimal model parameters for predicting the
55 presence of a lesion. The k NN uses a distance-based decision function for classifying lesions and MLP employs a nonlinear
56 activation function to distinguish lesions. The obtained accuracy from these two classifiers is less than 60%, thus more suitable
57 algorithms must be investigated to compare and establish the proposed work. Support vector machine has been investigated
58 using a linear and quadratic kernel, which is faster and has achieved optimal prediction outcomes for lesion classification.
59 These kernels are making SVM a powerful tool that can perform both linear and non-linear classification by mapping the inputs
60 to a high dimensional feature space and separates the categories by a gap that is as wide as possible. The ML outcomes are
61 evaluated through the results obtained from the radiologist’s report of the Perugia Hospital. These ML outcomes also have
62 been validated through established statistical measures. Preliminary results show the proposed method produces minimal
63 false-positives and false-negatives compared to other state-of-art methods and develop a viable anonymize method for mass
64 screening breast lesion detection in future.

65 Proposed Methodology

66 A “pipeline” schematic of the proposed work has been shown in Figure 1. At first a subject undergoing conventional screening
67 is asked to also undertake a parallel UWB imaging examination. In this case the conventional methods offered were echography,
68 mammography, magnetic resonance imaging. The radiologist in charge reviewed the conventional imaging data as usual to
69 make a decision regarding the screening outcome. The radiologist decisions have been considered as a gold standard identity
70 of each breast type investigated. Then, these gold standard labels of the breasts have been employed to train the supervised
71 machine learning algorithms to identify breast lesions automatically via the UWB imaging system. The outcomes of breast
72 lesion detection from radiologists and ML have been compared to ensure system performance. The details of intermediate
73 stages have been described in the following sections.

74 Apparatus description and set-up

75 All of the UWB imaging data used in this paper were collected on subjects using the microwave prototype (MammoWave, UBT
76 Srl, Italy) located at the Department of Diagnostic Imaging, Perugia Hospital, Perugia, Italy. All of the data was anonymized.
77 The functioning principle of the MammoWave system is based on the dielectric property difference between normal tissue and
78 tissue with lesions at microwave frequencies, i.e. the different behaviors that tissues display when irradiated by microwave
79 signals.

80 The hardware of the microwave prototype is composed of an aluminum cylindrical hub and shown in Figure 2. The
81 cylindrical hub represents a shield from external interferences and as a bearing structure for the entire device. On top of
82 the cylindrical hub lies the examination bed. The bed incorporates a plexiglass cup aimed at containing the breast of the
83 patient (facing down), with no pressure added to the breast tissue. Several sizes of this cup are available to accommodate the
84 examination of different breast sizes.

85 The transmit and receive (T_X and R_X respectively) antennas are positioned inside the hub and external to the cup, as shown
86 in Figure 3. They can rotate around the azimuth, to irradiate the breast (through T_X) and receive signals scattered by the breast




Figure 1. Flow diagram of the proposed work.

87 itself (through R_X). T_X and R_X have a distance from the center $a_1=20$ cm and $a_0=7$ cm, respectively. Both T_X and R_X are
 88 linearly polarized, operate in the 1-9 GHz frequency band and are connected to a vector network analyzer - VNA (Copper
 89 Mountain, Indianapolis, IN). Specifically, the received signals are the complex S21 data from the VNA. In particular, R_X can be
 90 rotated to measure the received signal at the points $rx_{np} \equiv (a_0, \phi_{np})$, displaced along a circular surface having radius a_0 , as
 91 shown in Figure 3.

$$E_{np,tx_m}(a_0, \phi_{np}; tx_m; f) = S21_{np,tx_m}(a_0, \phi_{np}; tx_m; f) \quad (1)$$

92 with $np = 1, \dots, N_{PT}$. The device uses M positions of the transmitting antennas, i.e. T_X can be rotated to transmit the signal
 93 from the points $tx_m \equiv (a_1, \phi_m)$, with $m = 1, 2, \dots, M$. Also, the device uses number of frequency (NF) samples in the band
 94 $B = [f_{min} \div f_{max}]$. The received signals are then processed through HP to calculate the field inside the cylinder; this field is then
 95 used to generate an image, which is a homogeneity map of dielectric properties¹⁹. Here, instead of using the received signals to
 96 generate images through the HP based algorithm, we employ ML methods on the raw signals to analyze and understand the
 97 difference between signals scattered from normal breast and breast with lesion to make decision about breast condition. The
 98 experiment was performed using Matlab R2017a tool in an Intel^R CoreTM i7 processor@ 3.60GHz based Windows 7 Enterprise
 99 64-bit operating system and it has 7856 MB NVIDIA Graphics Processing Unit (GPU).

100 In-vivo acquisition

101 Continuous clinical recruitment of volunteers is underway at Perugia Hospital, Italy. The clinical validation for the first forty-five
 102 volunteers has been approved by the Ethical Committee of Regione Umbria, Italy, (N. 6845/15/AV/DM of 14/10/2015). More
 103 recently, another partner hospital, Foligno Hospital, also in Italy, has been approved by the Ethical Committee of Regione
 104 Umbria, Italy, to join the clinical validation trials. This will extend the number of subjects to one hundred and fifty (N.
 105 10352/17/NCAV of 16/03/2017). This trial is a feasibility study of the method shown in Figures 2 and 3 employing microwave
 106 imaging to detect breast lesions, with the intention of gauging the potential of the proposed system for medical screening and
 107 localization of breast lesions. The list of a first pilot set of 18 subjects, who have been recruited under the aforementioned
 108 protocol and used for this study, is presented in Table 1. A previous table with 16 subjects was presented in ²⁸.



Figure 2. The UBT Microwave ‘MammoWave’ apparatus.




Figure 3. (a) Pictorial view of the system, where the transmit (T_X) and receive (R_X) antennas are placed inside the hub but external to the cup. (b) T_X (black dot) and R_X (red dot) can be moved around the azimuth, i.e. horizontal plane, on two circumferences having radius a_1 and a_0 , respectively²⁹.

109 All volunteers provided informed consent with five subjects undergoing the proposed microwave imaging for both breasts.
 110 Thirteen subjects underwent the prototype imaging as shown in Figure 3 for a single breast. This research was conducted
 111 adhering to the ethical standards of the institutional and/or national research committee incorporating the 1964 Helsinki
 112 declaration and its later amendments or analogous ethical standards. The study was performed in harmony with the Code of
 113 Ethics of the World Medical Association for experiments involving humans.

114 A conventional exam and full radiologist study was performed for each subject; echography and/or mammography or
 115 (limited to one case) magnetic resonance imaging were the conventional exams performed in these cases. The MyLab 70 xvg
 116 Ultrasound Scanner (Esaote, Italy) was the echographic method used; A Selenia LORAD Mammography System (Hologic,
 117 Marlborough, USA) was employed for mammography examinations; and a 3T scanner (Siemens Healthcare, Germany) was
 118 used for magnetic resonance imaging. The radiologists diagnosis is listed in Table 1 which presents the outcomes from the
 119 radiologists report along with the subject’s breast condition details. Where available, the breast type has been classified
 120 according to density, as defined by the American College of Radiology (ACR) scale ranging from ACR1 (extremely fatty breast)
 121 to ACR4 (extremely heterogeneous fibroglandular breast)³⁰. If present, the inclusion type was classified according to defined
 122 standards^{31–33}.

123 Following a volunteer’s agreement to participate, the clinical study coordinator supports each volunteer to place their breast
 124 correctly into the system cup. This cup is integrated into the prototype bed for improved comfort and patient stability. Exam
 125 data is processed via a computer interface and is observed by a system operator located in the room. Overall, the exam and data
 126 collection phase require approximately five minutes per breast. The transmit and receive antennas (T_X and R_X) are positioned
 127 on the azimuth plane at the same height which crosses the center of the breast of the subject being examined (after ensuring
 128 that the antennas half power beam angle include the breast). $M=15$ transmitting positions are employed, divided in 5 groups
 129 centered at 0° , 72° , 144° , 216° , and 288° on the azimuth plane; each group has 3 transmitting positions displaced from each
 130 other by 4.5° . For each transmitting position, four receiving positions, at 90° from each other are employed. The S21 was
 131 acquired at $NF=1601$ frequencies from 1 to 9 GHz in $\Delta f = 5$ MHz increments for each T_X and R_X position. The received
 132 signals are processed as follows:

Table 1. Subject lists; details and related radiologist review. Overall shows 12 healthy and 11 non-healthy tissues. Included in the non-healthy is one post-surgical breast with seroma.

Subject index and breast (left/ right)	Year of birth	Breast type	Diagnostic test	Output of the radiologist study
01R	1983	ACR4	ecography	Healty
01L	1983	ACR4	ecography	Healty
02R	1936	ACR2	mammography	carcinoma papillary
03R	1960	N/A	magnetic resonance	carcinoma infiltrating grade 2
04R	1987	ACR4	ecography	Healty
04L	1987	ACR4	ecography	Healty
05L	1987	N/A	ecography	benign fibroadenoma
06L	1975	ACR2	ecography + mammography	benign fibroadenoma
07R	1980	ACR3	ecography + mammography	benign microcalcifications
08R	1929	ACR4	ecography + mammography	carcinoma
09R	1963	ACR3	mammography	Healthy
09L	1963	ACR3	mammography	benign fibroadenoma
10L	1964	ACR4	mammography	Healthy
10R	1964	ACR4	mammography	Healthy
11R	1946	ACR2	mammography	Healthy
12R	1966	ACR3	mammography	carcinoma (4 cm), b5
13L	1971	ACR3	mammography	carcinoma
14L	1996	N/A	ecography	benign fibroadenoma
15R	1969	ACR4	mammography	microcalcifications
16R	1948	ACR2	mammography	Healthy
17R	1971	ACR4	mammography	Healthy
17L	1971	ACR4	mammography	Healthy
18R	1983	ACR3	mammography	Healthy

1. Let us consider the first transmitting antennas group and the antennas in the group are numbered as tx_1, tx_2, tx_3 ; the first transmitting group is assumed to be centered at $\phi = 0^\circ$. For each transmitting antenna of the group, we consider the signals received at $N_{PT} = 4$ points displaced at $\phi = 45^\circ, 135^\circ, 225^\circ, 315^\circ$ with respect to the corresponding transmitting antenna.

2. To remove skin artefacts, we generate the following signals:

$$E_{1-2}(a_0, \phi_{np}; tx_{1-2}; f) = E_{np,tx_m}(a_0, \phi_{np}; tx_1; f) - E_{np,tx_m}(a_0, \phi_{np}; tx_2; f) \quad (2)$$

$$E_{2-3}(a_0, \phi_{np}; tx_{2-3}; f) = E_{np,tx_m}(a_0, \phi_{np}; tx_2; f) - E_{np,tx_m}(a_0, \phi_{np}; tx_3; f) \quad (3)$$

where, $\phi_{np} = 45^\circ, 135^\circ, 225^\circ, 315^\circ$

3. The procedure is repeated for the other four transmitting groups. It follows that, for each microwave exam, we have 40 signals in the frequency domain.

Supervised Machine Learning

The data gathered from the microwave examination are considered for classification purpose and/or to predict future lesion detection instances. This is the first investigation to use microwave clinical data from this apparatus undergoing machine learning classification. Labeled information about the healthy and non-healthy breast pattern have been gathered from the output of the radiologist study. Specifically, a healthy breast is a breast with no lesion, while a non-healthy breast is a breast containing a lesion which may be benign or malignant. The accumulated clinical data have high variance in Euclidean space, so both linear and non-linear classifiers are employed to optimised performance for identifying healthy and non-healthy subjects. There are several ML algorithms present for classification tasks in the literature, some of them being very problem specific while others aim to be more general requiring an investigation approach to be fitted to available data. The selection of appropriate ML methods is quite intuitive and the data distribution in the plane needs to be initially observed. It has been found that the gathered

151 microwave data are non-linearly distributed in the plane. Thus, the leading supervised and non-linear classifiers, KNNs, MLPs,
 152 SVMs have been examined, where cross validation techniques, augmented by random sub-sampling methods are employed to
 153 assess the performance, through statistical metrics, and discover the most appropriate classification model. Initial results from
 154 non-linear classifiers such as, KNNs and MLPs were unsatisfactory. The radial basis function (RBF) could be an option, but it
 155 performs well where the data are in loop or spherical shape and circular decision boundary can differentiate the groups. Though,
 156 the data are non-linear they are not spherically distributed, hence the SVM has been implemented with a linear kernel function
 157 because, although the data distribution appears non-linearly separable in the 2D plane, there is still a possibility to classify
 158 most lesion instances accurately by linear decision boundary while the data are being projected into hyperplane i.e., impossible
 159 to visualise. The linear kernel of SVM also failed to achieve satisfied performance, and SVM has been experimented with
 160 quadratic kernel function which outperformed other classification techniques.

161 Cross validation and performance evaluation

162 A cross validation technique has been used to assess, enhance predictive outcomes, and select models for developed ML
 163 prototypes. This has been done by repeated random sub-sampling of the data, which is also known as Monte Carlo cross-
 164 validation³⁴. The dataset has been randomly partitioned to select the training and validation dataset, where training and
 165 validation sets have been used to train and evaluate performance of a selected ML model. The ratio of training and testing data
 166 has been specified during each round e.g., training has been started with 10% randomly selected data when rest of the 90%
 167 data have been considered as validation/testing data. The amount of training data has been increased by 10% while amount of
 168 validation dataset decreased by 10% in each round and this process has been repeated till the model has not overfitted. Each
 169 model has been run 25 rounds to select the appropriate ratio of training and testing and found 40% of training and 60% of
 170 testing data is necessary to prevent the ML algorithms from overfitting. The results (statistical metrics) have been aggregated
 171 and averaged over all the rounds. A number of statistical metrics³⁵, accuracy, true positive rate (TPR) or sensitivity, true
 172 negative rate (TNR) or specificity, positive predictive value (PPV), and negative predictive value (NPV) have been used to
 173 investigate the classification performance of the classifiers. A receiver operating characteristic (ROC) curve has been generated
 174 for each ML model with the validation dataset to illustrate the diagnostic ability and stability of the classification system with
 175 different discrimination threshold. Subsequently, Matthews Correlation Coefficient (MCC)³⁶ and Youden's index³⁷ have been
 176 implemented to investigate the classification outcomes, where, MCC and Youden's index estimate quality of classification and
 177 probability of the informed decision respectively. The outcomes and its analysis have been described in next section.

178 Results Analysis

179 According to the radiologist's review, 12 healthy breasts and 11 non-healthy breasts, i.e. breasts with lesions, underwent the
 180 microwave exam of the 18 subjects. As described in the previous section, each microwave exam leads to 40 different patterns in
 181 the frequency domain. As an example, Figure 4 below shows $E_{1-2}(a_0, \phi_{np}; tx_{1-2}; f)$ for $\phi_{np} = 45^\circ$ for one healthy and one
 182 non-healthy breast.




Figure 4. Example signals of healthy and non-healthy patterns in 45° .

183 *k*-nearest neighbor classifier

184 Initially, the investigation began by employing the *k*-NN classifier³⁸. The classifier is particularly simple, measuring the
 185 proximity of features in the hyperspace without any assumption of the underlying data distribution to predict a category, making
 186 it flexible for decision making. Two effective distance metrics, Euclidean and Mahalanobis perform well with *k*-NN, but
 187 Mahalanobis distance requires the inversion of covariance matrix which could increase the computational overhead. Therefore,
 188 the Euclidean distance is considered here to measure the distance of a feature vector from its nearest neighbor. The *k* is chosen
 189 as odd for this two-class problem that one pattern could not predict under the same class label by the classifier. Table 2 displays
 190 the classification outcomes, where *k* is varied from 1 to 5 and 10%, 20%, 30%, and 40% data are randomly selected for
 191 the training phase. The results show that the algorithm exhibited good performance with increasing training data volume as
 192 expected. Here, *k* = 1, produces the optimal result among other NNs with 40% of training data volume. It attained the testing
 193 accuracy 0.608 (≡60.8%). TPR or sensitivity measures the ability of the algorithm to identify the non-healthy subjects, which
 194 is 0.541(≡54.1%) in the case of *k* = 1. It could correctly identify the subjects with lesions with a rate of 0.667 (≡66.7%).
 195 PPV and NPV are influenced by the prevalence of having a lesion in the breast that is being tested. In case of *k* = 1, PPV
 196 and NPV the probability that the subjects with positive lesion identification truly have the lesion and negative identification
 197 of lesions truly do not have the lesion. The 1NN produces fewer false predictions close to the decision boundary bringing
 198 improved accuracy over 3NN and 5NN; also, truly positive prediction for having lesion and vice versa. Additionally, the MCC
 199 measurements over prediction results are also not cogent for 1NN, 3NN, and 5NN. The average MCC is approximately 0.206
 200 in case of 1NN and decreases towards 0 with the increment of *k*. This trend states that the addition of random predictions with
 201 greater number of NN. The average proportions obtained from Youden's index are also very low, 0.206, 0.130, and 0.086 for
 202 1NN, 3NN, and 5NN respectively further indicating the probability to predict those lesions is random and unreliable. This
 203 index works along with the ROC curve, the outcomes have been correlated at the discussion of ROC analysis. Therefore, the
 204 overall performance of 1NN is better than the other NNs because of data compactness, where one nearest neighbor results in a
 205 good prediction if a greater number of neighbors have been chosen, the misclassification increases.

Table 2. Results obtained from nearest neighbor algorithm.

NNs	% of Training Data	Accuracy	Sensitivity	Specificity	PPV	NPV	MCC	Youden's Index
1NN	10	0.555	0.567	0.545	0.521	0.590	0.203	0.202
	20	0.551	0.441	0.650	0.532	0.563	0.225	0.225
	30	0.586	0.580	0.592	0.551	0.620	0.186	0.186
	40	0.608	0.541	0.667	0.587	0.624	0.210	0.209
3NN	10	0.519	0.648	0.405	0.488	0.569	0.143	0.142
	20	0.528	0.448	0.600	0.505	0.544	0.125	0.124
	30	0.559	0.460	0.652	0.553	0.562	0.123	0.122
	40	0.568	0.520	0.609	0.531	0.598	0.134	0.132
5NN	10	0.532	0.434	0.619	0.499	0.555	0.084	0.082
	20	0.541	0.412	0.656	0.517	0.555	0.088	0.087
	30	0.549	0.577	0.524	0.513	0.588	0.092	0.089
	40	0.550	0.419	0.665	0.522	0.567	0.082	0.081

206 Multilayer perceptron classifier

207 Two different multilayer perceptron^{39,40} algorithms are studied where, each algorithm is created with one hidden layer and
 208 the number of nodes in the hidden layer are decided using a 'rule of thumb' ($\sqrt{(number\ of\ inputs + number\ of\ outputs)} +$
 209 $(a\ constant\ between\ 1\ to\ 10\ set\ by\ experimentally)$). The optimal size of the hidden layer is decided typically between the
 210 size of the input and output layers. The bias and weights are initialized randomly, the learning rate $\eta = 0.1$ is varied up to 0.99.
 211 The output of the layers is determined by the hyperbolic tangent sigmoid transfer function. The Mean Square Error (MSE)
 212 $= \frac{1}{N} \sum_{i=1}^N (t_i - a_i)^2$ is calculated for each output to back-propagate and update the weights, where *t* and *a* signify the targets and
 213 outputs, respectively.

214 First, the network is trained using the Levenberg-Marquardt (LM) algorithm which adaptively varies the parameter updates
 215 and performs better (because of the weight updation using a damping coefficient) than the simple gradient decent method
 216 that defines simple first order iterative optimization function and finds the local minimum, local maximum for parameter
 217 updating. The training stops when the maximum number (=1000) of epochs is reached, where one set of weight updating using
 218 backpropagation is considered as one epoch. Table 3 presents the results for both MLP using Levenberg-Marquardt MLP_{LM} ⁴¹
 219 and Bayesian-Regularization (BR) backpropagation MLP_{BR} ⁴². In the case of MLP_{LM} , the testing accuracy increases for up
 220 to 40% training data, but the increment rate is negligible, and reached 0.532 (≡53.2%), but results 0.076 (≡7.6%) sensitivity
 221 or TPR indicates the network can only identify 7.6% subjects correctly. However, it shows good performance in recognizing

222 subjects without lesions (TNR=0.936). Also, the probability of the prediction for identification of subjects having or not having
 223 lesions is slightly more than 50%. In the case of MLP_{BR} , the overall performance is similar to MLP_{LM} . The maximum testing
 224 accuracy reached 0.538 ($\equiv 53.8\%$) when 40% training data is used but results only 0.038 TPR demonstrates the inability to
 225 make predictions about lesions whereas 0.951 specificity or TNR shows a strong performance in predicting the absence of a
 226 lesion. Therefore, neither MLPs could predict the healthy breast pattern, but did make acceptable predictions for subjects with
 227 lesions. The probability of the prediction for identification of subject with or without lesions is between 0.511 to 0.542 for both
 228 MLP_{BR} and MLP_{LM} . Additionally, the estimation of MCC and Youden's statistic state insignificant power of MLPs to identify
 229 breast lesions. MCC of MLP_{LM} and MLP_{BR} are 0.019 and 0.082 respectively, implies the performance of MLPs are no better
 230 than random prediction with a large, unacceptable, misclassification rates. Subsequently, Youden's statistics of MLP_{LM} and
 231 MLP_{BR} are 0.018 and 0.082 respectively. The zero tendency of the indices show the high proportion of false positives and false
 232 negatives. Misidentifications have been found here because the error surfaces are very complex for both of these networks and
 233 have stagnant into several local minima, producing unexpected outcomes for healthy and non-healthy patient identification.

Table 3. Results obtained from multilayer perceptron algorithm.

MLPs	% of Training Data	Accuracy	Sensitivity	Specificity	PPV	NPV	MCC	Youden's Index
MLP_{LM}	10	0.498	0.111	0.847	0.394	0.514	0.020	0.019
	20	0.530	0.052	0.962	0.551	0.529	0.010	0.009
	30	0.531	0.190	0.825	0.482	0.542	0.016	0.015
	40	0.532	0.076	0.936	0.511	0.534	0.031	0.030
MLP_{BR}	10	0.532	0.035	0.953	0.428	0.532	0.081	0.081
	20	0.533	0.032	0.934	0.462	0.513	0.088	0.088
	30	0.538	0.037	0.957	0.414	0.527	0.082	0.082
	40	0.538	0.038	0.951	0.425	0.538	0.078	0.078

234 Support vector machine classifier

235 Subsequently, the SVM is investigated with two different kernel functions to acquire the hyperplane that can separate
 236 healthy and non-healthy subjects. Table 4 shows the results for classification of the 2 subject types where, SVM_L and SVM_Q
 237 represents the SVMs using the linear and quadratic kernel functions^{43,44} for prediction. SVM_L uses the optimization method,
 238 $c = \sum_i w_i k(s_i, x) + b$ where, subject pattern vector x is targeted to classify, s_i is the support vector, w_i is weight, and b is the
 239 bias. Here, the linear kernel function is k . The vector x is considered a member of the lesion free group when, $c \geq 0$ or lesion
 240 group otherwise. This creates a hyperplane that achieved better accuracy than the other classifiers used above. SVM_L is trained
 241 using 10% to 40% data and associated testing results are shown in the table. It produces the highest testing accuracy with 40%
 242 training data which is 0.620 ($\equiv 62.0\%$). It achieved TNR of 0.998 ($\equiv 99.8\%$) in that case, which indicates a good performance
 243 to identify the subjects with no lesions, but TPR (maximum 0.193 $\equiv 19.3\%$ among all cases of SVM_L) shows a very weak
 244 performance in identifying subjects with lesions. Though the probability in identifying subjects who truly have lesions is better
 245 than the subjects who truly do not have lesions. The number of false negatives are continuously high, but false positives are low,
 246 which resulting an average MCC 0.319 for SVM_L . Though, MCC is better than the other algorithms, it is still not powerful
 247 enough to reduces false negatives. The average Youden's index is 0.180 which is close to the Youden's index obtained from
 248 1NN because, SVM_L produced more false negatives and fewer false positives, whereas 1NN resulted in fewer false negatives
 249 and more false positives, thus the total number false predictions are high in both the cases.

Table 4. Results obtained from SVM using different kernel.

SVMs	% of Training Data	Accuracy	Sensitivity	Specificity	PPV	NPV	MCC	Youden's Index
SVM_L	10	0.620	0.191	0.997	0.981	0.584	0.319	0.182
	20	0.619	0.184	0.998	0.985	0.584	0.318	0.178
	30	0.616	0.194	0.996	0.977	0.579	.324	0.180
	40	0.620	0.171	0.998	0.989	0.589	0.318	0.181
SVM_Q	10	0.985	0.969	0.996	0.985	0.973	0.959	0.956
	20	0.984	0.967	0.997	0.983	0.972	0.963	0.960
	30	0.984	0.965	0.996	0.985	0.970	0.963	0.959
	40	0.989	0.977	0.997	0.985	0.981	0.955	0.951

250 Subsequently, SVM_Q have been employed to obtain an improved testing accuracy to differentiate subjects by minimising the
 251 gap between the two groups. The considered quadratic function is $\min_x \frac{1}{2}x^T Hx + c^T x$, where $Ax \leq b$, c is a real valued vector,
 252 H is real symmetric matrix, A is real matrix, b is a real vector, and the notation $Ax \leq b$ means that every entry of the vector Ax is
 253 less than or equal to the corresponding entry of the vector b . The quadratic programming aims to find the vector x which could

254 minimize that function. This function creates the best hyperplane to classify the subjects here. SVM_Q achieved 0.989 ($\equiv 98.9\%$)
 255 testing accuracy to identify the lesion affected and unaffected subjects. Correspondingly, TPR (is 0.9770 $\equiv 97.70\%$) and TNR
 256 (is 0.997 $\equiv 99.7\%$) both are high in this case, which indicates a good implementation of the hyperplane for separation which
 257 could correctly identify those subjects with and without lesions. Also, the high probabilities (PPV and NPV) support the results.
 258 In addition, MCC and Youden's index both are high (i.e., 0.960 and 0.956 respectively) in this case. Figure 5 illustrates the
 259 outcomes more clearly, where no false positives, and few false negatives have been found in each run. This greatly influences
 260 the score of MCC and Youden's index and proves the strong ability of the SVM_Q model to identify breast lesions.

		Predicted	
		Class 1	Class 2
Actual	Class 1	776	25
	Class 2	0	897

(a)

		Predicted	
		Class 1	Class 2
Actual	Class 1	697	24
	Class 2	0	829

(b)

		Predicted	
		Class 1	Class 2
Actual	Class 1	644	23
	Class 2	0	744

(c)

		Predicted	
		Class 1	Class 2
Actual	Class 1	553	13
	Class 2	0	670

(d)

Figure 5. Confusion Matrices of SVM_Q : (a) 10% training data, (b) 20% training data, (c) 30% training data, and (d) 40% training data are used.

261 Figure 5 shows the outcomes of SVM_Q more closely from the confusion matrices. The numeric values listed in the confusion
 262 matrix with a blue background demonstrates the correct classification of lesion affected and unaffected patterns. Very few
 263 misclassification occurred here (between 1.1% to 1.5%). Without lesion and with lesions are denoted by class-1 and class-2,
 264 respectively. It is shown that all the normal tissue subjects are classified correctly, few misclassifications are found with the
 265 training data variation (10% to 40%) in each case of SVM_Q . The most important part of this lesion classification is to reduce
 266 the negative predictions (including false positive and false negative). It is found from the confusion matrices that the false
 267 prediction rate of lesion detection is zero in each case (all the patterns of class-2 are being predicted as class-2). False positive
 268 rate (FPR) is zero (from Figure 5a to 5d) for all the cases of SVM_Q , which implies all the non-lesion breast patterns are correctly
 269 classified but, few false negative rates (FNR) occurred, 0.027 (2.7% for 10% training data), 0.028 (2.8% for 20% training data),
 270 0.029 (2.9% for 30% training data), and 0.019 (1.9% for 40% training data). Results here are far exceed other state-of-art
 271 methods reported to miss 15% of lesions. The effect of this classification and misclassification has been discussed earlier.




Figure 6. (a) Actual data distribution in two-dimensional plane, (b) decision boundary produced by 1NN classifier, (c) decision boundary produced by SVM with quadratic kernel.

272 As reported, SVM_Q has achieved the best performance among the classifiers investigated here in terms of accuracy,
 273 sensitivity, and specificity. It's important here to visualize the data with corresponding hyperplanes formed by the ML
 274 algorithms. Figure 6 presents the data distribution along with the two optimal decision boundaries produced by 1NN and
 275 SVM_Q . Figure 6a represents the breast with lesion and healthy breast data distribution in a two-dimensional feature space for
 276 visualization of the studied data pattern for the classification purpose. All the healthy breast patterns scatter the microwaves in

277 a particular way whereas a breast lesion may occur in any position of the breast, which produces different scattering effect
 278 for the breasts with lesion. Hence, it has been found healthy data formed the cluster in the center and lesion data scattered
 279 around that group. Thus, these two groups of data could only be handled by the parabolic or quadratic curve, which has been
 280 made by SVM_Q here (decision boundary of Figure 6c). In addition, Figure 6b shows the decision boundary to correlate the
 281 misidentifications obtained by 1NN algorithm, which also have been illustrated in Table 4.

282 Finally, the obtained classification results are compared to conclude the investigation. Figure 7 shows the visual comparison
 283 of average classification accuracy, sensitivity, specificity, MCC, and Youden's index to make the contrast over performance,
 284 where x and y -axis represent different classifiers and accuracy, sensitivity, specificity, MCC, and Youden's index respectively. It
 285 can be seen, the performance of SVM_Q is better than other classifiers attempted here in terms of average accuracy (98.55%),
 286 sensitivity (96.95%), specificity (99.65%), MCC (96%), and Youden's statistic (95.6%) which illustrates robust ability to detect
 287 breast lesions from new microwave data.




Figure 7. Comparison of performance metrics for all classifiers, (a) average accuracy, (b) average sensitivity, (c) average specificity, (d) average MCC, (e) Youden's index.

288 The ROC curve has also been plotted and shown in Figure 8 to compare and analyse the diagnostic ability of all three
 289 classifiers, where the number of nearest neighbour, learning rate, and threshold for detection have been varied for NNs, MLPs,
 290 and SVMs respectively. The area under curve (AUC) has also been determined for each classifier. The x and y -axes of
 291 Figures 8a, 8b, 8c represent false positive rate (FPR) or (1-specificity) and true positive rate (TPR) or sensitivity respectively.
 292 Though, the accuracy of 1NN is better than other classifiers, except SVM_L and SVM_Q , the AUC of overall KNN (Figure 8a) is
 293 only 0.599 indicating the large presence of false predictions. In the case of MLP_{BR} and MLP_{LM} , AUCs are only 0.389 and 0.446
 294 respectively. Both the MLPs produce vast amount of false predictions, as discussed earlier, and generate low AUC, and the
 295 tendency of performance is highly random in terms of all performance metrics. Though, the accuracy of SVM_L is approximately
 296 0.619, it made a large number of false predictions and produced an AUC of 0.228 when the threshold had been varied. The
 297 highest AUC ($\equiv 0.937$) is created by SVM_Q as this delivers the lowest number of false predictions among all tested and scored
 298 the highest performance metric for all the cases.

299 The parametric (i.e., MLP, and SVM) and non-parametric (i.e., KNN) both types of learning techniques have been used here
 300 to obtain optimal performance on the currently available dataset. The work has two main limitations, high dimensional feature
 301 space and availability of data (i.e., still ongoing). The dimension of the feature space has not been reduced in this study and
 302 considered as future scope. The KNN and SVM make the decision based on the similarity measure, whereas MLP depends upon
 303 features. The non-parametric KNN does not require any assumption of feature distribution to check similarity, but it requires a
 304 large amount of uncorrelated and independent training data in order to make good predictions, whereas the data used in this




Figure 8. ROC curve analysis of the classifiers, (a) KNN, (b) MLP using Levenberg-Marquardt and Bayesian-Regularization backpropagation functions, (c) SVM using linear and quadratic kernels.

305 study are highly correlated and high dimensional in nature. Thus, the KNN has overfitted with 40% of training data, the overall
 306 performance is unsatisfactory and also deteriorated with the increment of k . Both the MLPs MLP_{BR} and MLP_{LM} are parametric
 307 in nature and make assumptions of feature space to minimise cost function and get optimised weights. Though, the function
 308 LM which directs MLP_{LM} is well known for optimising cost function but, choice of damping parameter played vital role for the
 309 study and the model stuck in local minima within 50 epochs which prevents the model to optimise weights and create good
 310 decision boundary. Thus, the damping parameter needs further tune to obtain good results. The Bayesian-Regularization (BR)
 311 is a well known optimisation technique that works well with MLP even if the data are high dimensional. But, the possible
 312 reason for incorrect predictions of MLP_{BR} is miss-specification of the model which indicates the function model does not suit
 313 BR for this problem. The SVM has been used by two different kernel functions, linear (SVM_L) and quadratic (SVM_Q). The
 314 methods rely on similarity, unlike the KNN model, SVMs is sensitive to the curse of dimensionality problem while the the
 315 features are not engineered to uncorrelated values. SVM_L is well known for its ability to separate non-linear data linearly in the
 316 higher feature space. Here, the dot product weight and features with the conventional bias have been employed to create a
 317 linear hyperplane and maximise the gap between support vectors and samples. Also, the estimation of bias is trivial in this case,
 318 thus, the maximisation of the linear kernel function created hard margins and increased errors as a result. On the other hand,
 319 SVM_Q is an extended version (2^{nd} order polynomial) of SVM_L which creates a soft margin in the feature space. The SVM_Q
 320 is focused on the minimisation problem. The kernel has two other variables, c and x which denote penalty constraint and the slack
 321 variable respectively. This is the advantage found using SVM_Q , where x minimise error and c minimise the gap at the same time
 322 so that the non-linear boundary has been formed for one group (healthy or non-healthy) and rest of the samples fall in the other
 323 side of the boundary.

324 Discussion

325 The frequency domain scattering parameters of subject breasts are obtained using the UWB microwave mammogram apparatus
 326 described above in an ongoing clinical trial. The same subject breasts included in this study have also undergone radiologist
 327 scrutiny obtained using conventional imaging methods (echography and/or mammography or (limited to one case) magnetic
 328 resonance imaging), which has been used as labeled information. The microwave mammogram apparatus clinical data are
 329 pre-processed and transformed for machine learning approaches. Various algorithms are trained and tested to differentiate
 330 lesion-containing and lesion-free breast tissues. Here, breasts with lesions may be benign or malignant. The experimental
 331 results show that the quadratic kernel of SVM has successfully created the hyperplane and maximizes the margins between the
 332 support vectors, resulting in a sensitivity for breast lesion classification equal to 97%. Such value outperforms the sensitivity
 333 given in¹⁶ and¹⁷, which are 74% and 90% respectively, where machine learning is not employed. The successful employment of
 334 machine learning on clinical data obtained using a microwave mammogram could help the radiologist in the diagnosis process.
 335 The integrated system with microwave non-ionizing imaging augmented by machine learning algorithms can be a step change
 336 in mass breast screening deployment. The system could be deployed across all female age groups, and during pregnancies, in
 337 more local settings, increasing the detection and hence survival rates of breast cancer sufferers.

338 This study aimed to differentiate between breast with or without lesions, but the type of lesion cannot yet be identified at
 339 this stage. The authors are currently gathering more clinical data to understand the category and property of lesions through
 340 the MammoWave device. Once more data have been gathered to make generalized decisions about lesions, this work will

341 be extended further to identify the type of lesions automatically using ML. Although, the SVM_Q has worked very well in
342 terms of all statistical performance metrics, the study is limited by some factors and those are considered for future work. The
343 data used here have a high dimension and values are correlates, which placed the experimented ML methods in the curse
344 of dimensionality problem. Thus, suitable dimensionality reduction techniques will be investigated in the next stage. The
345 analysis of ROC curve shows that performance improvement for all experimented ML algorithms are possible. This study is
346 an empirical study on ML application of automatic lesion detection to investigate suitable classifiers to categorise the data in
347 hyperspace. The performance analysis directs focus on hyperplanes created by a quadratic function, thus the fine tuning of
348 the parameters such as, slack variable and penalty parameter would be observed subsequently. Also, the parameter bias of
349 other algorithms would be taken into account since the performance of current prototype may vary when the type of lesion will
350 be identified with new datasets. Expert clinical input will be ensured in further research to meet clinical expectations in the
351 assistance of breast lesion identification and classification. In addition, deep learning approaches will be investigated to provide
352 higher sensitivity and specificity towards automation.

353 References

- 354 1. Petracci, E. et al. Risk factor modification and projections of absolute breast cancer risk. *J. Natl. Cancer Inst.* 103, 1037–1048
355 (2011).
- 356 2. <https://www.wcrf.org/int/cancer-facts-figures/data-specific-cancers/breast-cancer-statistics>.
- 357 3. <http://ec.europa.eu/eurostat/statisticsexplained/index.php/Being-young-in-Europe-today-demographic-trends>.
- 358 4. Sardanelli, F. et al. Mammography: an update of the eusobi recommendations on information for women. *Insights into*
359 *imaging* 8, 11–18 (2017).
- 360 5. Miglioretti, D. L. et al. Radiation-induced breast cancer incidence and mortality from digital mammography screening: a
361 modeling study. *Annals internal medicine* 164, 205–214 (2016).
- 362 6. Fuller, M. S., Lee, C. I. and Elmore, J. G. Breast cancer screening: an evidence-based update. *The Med. clinics North Am.*
363 99, 451 (2015).
- 364 7. Council, N. R. et al. Mammography and beyond: developing technologies for the early detection of breast cancer (National
365 Academies Press, 2001).
- 366 8. Rahman, A. et al. Electromagnetic performances analysis of an ultra-wideband and flexible material antenna in microwave
367 breast imaging: To implement a wearable medical bra. *Sci. reports* 6, 38906 (2016).
- 368 9. Li, X. and Hagness, S. C. A confocal microwave imaging algorithm for breast cancer detection. *IEEE Microw. Wireless*
369 *components letters* 11, 130–132 (2001).
- 370 10. Bond, E. J. et al. Microwave imaging via space-time beamforming for early detection of breast cancer. *IEEE Transactions*
371 *on Antennas Propag.* 51, 1690–1705 (2003).
- 372 11. Nikolova, N. K. Microwave imaging for breast cancer. *IEEE microwave magazine* 12, 78–94 (2011).
- 373 12. Lazebnik, M. et al. A large-scale study of the ultrawideband microwave dielectric properties of normal breast tissue
374 obtained from reduction surgeries. *Phys. Medicine and Biol.* 52, 2637 (2007).
- 375 13. Lazebnik, M. et al. A large-scale study of the ultrawideband microwave dielectric properties of normal, benign and
376 malignant breast tissues obtained from cancer surgeries. *Phys. Medicine and Biol.* 52, 6093 (2007).
- 377 14. Conceição, R. C., Mohr, J. J. and O'Halloran, M. An introduction to microwave imaging for breast cancer detection
378 (Springer, 2016).
- 379 15. Meaney, P. M. et al. Initial clinical experience with microwave breast imaging in women with normal mammography. *Acad.*
380 *radiology* 14, 207–218 (2007).
- 381 16. Preece, A. W., Craddock, I., Shere, M., Jones, L. and Winton, H. L. Maria m4: clinical evaluation of a prototype
382 ultrawideband radar scanner for breast cancer detection. *J. Med. Imaging* 3, 033502 (2016).
- 383 17. Tiberi, G. and Raspa, G. Apparatus for testing the integrity of mammary tissues. EU patent no.0001413526 (2012).
- 384 18. Ghavami, N. et al. Uwb microwave imaging of objects with canonical shape. *IEEE Transactions on Antennas Propag.* 60,
385 231–239 (2012).
- 386 19. Tiberi, G. et al. Sensitivity assessment of a microwave apparatus for breast cancer detection, european congress of radiology.
387 ECR 2018, C-1390, 10.1594/ecr2018/C-1390 (2018).

- 388 **20.** Yassin, 393 N. I. et al. Machine learning techniques for breast cancer computer aided diagnosis using different image
389 modalities: A systematic review. *Comput. Methods Programs Biomed.* (2017).
- 390 **21.** Yamamoto, Y. et al. Quantitative diagnosis of breast tumors by morphometric classification of microenvironmental
391 myoepithelial cells using a machine learning approach. *Sci. Reports* 7, 46732 (2017).
- 392 **22.** Hosny, A., Parmar, C., Quackenbush, J., Schwartz, L. H. & Aerts, H. J. Artificial intelligence in radiology. *Nat. Rev. Cancer*
393 1 (2018).
- 394 **23.** Davis, S. K., Van Veen, B. D., Hagness, S. C. and Kelcz, F. Breast tumor characterization based on ultrawideband
395 microwave backscatter. *IEEE transactions on biomedical engineering* 55, 237–246 (2008).
- 396 **24.** Song, H., Li, Y., Coates, M. and Men, A. Microwave breast cancer detection using empirical mode decomposition features.
397 arXiv preprint arXiv:1702.07608 (2017).
- 398 **25.** Sekkal, W., Merad, L. and Meriah, S. M. A comparative study for breast cancer detection by neural approach for different
399 configurations of the microwave imaging system. *Prog. In Electromagn. Res.* 65, 69–78 (2018).
- 400 **26.** Li, Y., Porter, E., Santorelli, A., Popovic', M. and Coates, M. Microwave breast cancer detection via cost-sensitive ensemble
401 classifiers: Phantom and patient investigation. *Biomed. Signal Process. Control.* 31, 366–376 (2017).
- 402 **27.** Gerazov, B. and Conceicao, R. C. Deep learning for tumour classification in homogeneous breast tissue in medical
403 microwave imaging. In *Smart Technologies, IEEE EUROCON 2017-17th International Conference on*, 564–569 (IEEE,
404 2017).
- 405 **28.** Sani, L., et al. "Microwave apparatus for testing breast integrity based on Huygens principle: clinical validation on 16
406 subjects.", *Loughborough Antennas & Propagation Conference (LAPC 2017)*, page (5 pp.) (IET, 2017).
- 407 **29.** Sani, L., et al. "Novel microwave apparatus for breast lesions detection: Preliminary clinical results.", *Biomedical Signal*
408 *Processing and Control* 52 (2019): 257-263.
- 409 **30.** American College of Radiology and others, A. C. et al. *Acr practice guideline for the performance of screening and*
410 *diagnostic mammography. Pract. guidelines technical standards.* Reston, Va: Am. Coll. Radiol. 525–34 (2008).
- 411 **31.** Tavassoli, F. A. and Devilee, P. *Pathology and genetics of tumours of the breast and female genital organs* (Iarc, 2003).
- 412 **32.** Lakhani, S. et al. *WHO classification of tumours of the breast, vol. 4, chap. Lobular neoplasia, 77–80* (IARC Press, Lyon,
413 France, 2012).
- 414 **33.** Perry, N. and Puthaar, E. *European guidelines for quality assurance in breast cancer screening and diagnosis* (European
415 Communities, 2006).
- 416 **34.** Picard, R. R. and Cook, R. D. Cross-validation of regression models. *J. Am. Stat. Assoc.* 79, 575–583 (1984).
- 417 **35.** Powers, D. M. *Evaluation:from precision, recall and F-measure to ROC, informedness, markedness and correlation* (Bioinfo
418 Publications, 2011).
- 419 **36.** Matthews, B. W. Comparison of the predicted and observed secondary structure of t4 phage lysozyme. *Biochimica et*
420 *Biophys. Acta (BBA)-Protein Struct.* 405, 442–451 (1975).
- 421 **37.** Youden, W. J. Index for rating diagnostic tests. *Cancer* 3, 32–35 (1950).
- 422 **38.** Cover, T. and Hart, P. Nearest neighbor pattern classification. *IEEE transactions on information theory* 13, 21–27 (1967).
- 423 **39.** Rosenblatt, F. *Principles of neurodynamics. perceptrons and the theory of brain mechanisms.* Tech. Rep., CORNELL
424 AERONAUTICAL LAB INC BUFFALO NY (1961).
- 425 **40.** Rumelhart, D. E., McClelland, J. L. and PDP Research Group, C. (eds.) *Parallel Distributed Processing: Explorations in*
426 *the Microstructure of Cognition, Vol. 1: Foundations* (MIT Press, Cambridge, MA, USA, 1986).
- 427 **41.** Levenberg–marquardt methods with strong local convergence properties for solving nonlinear equations with convex
428 constraints. *J. Comput. Appl. Math.* 172, 375 – 397 (2004). DOI <https://doi.org/10.1016/j.cam.2004.02.013>.
- 429 **42.** Burden, F. and Winkler, D. Bayesian regularization of neural networks. In *Artificial neural networks*, 23–42 (Springer,
430 2008).
- 431 **43.** Abe, S. *Support vector machines for pattern classification*, vol. 2 (Springer, 2005).
- 432 **44.** Smola, A. J. and Scholkopf, B. *Learning with kernels*, vol. 4 (Citeseer, 1998).

433 **Author contributions statement**

434 G.T., L.S., A.V., and G.R. have designed and manufactured the UWB microwave non-ionising apparatus with associated signal
435 processing techniques comprising MammoWave. Subjects have been recruited and screened through MammoWave. M.D. has
436 conducted the conventional radiological investigation on the same subjects and confirmed the outcomes of UWB MammoWave
437 experiment. S.P.R. and M.D. have performed the machine learning algorithms and analysed the prediction outcomes for
438 automatic breast lesion detection through clinical UWB MammoWave. S.D. and M.G. have supervised the work. S.D. managed
439 the experiments performed along with co-authors at LSBU and instigated the collaborative work on this paper between teams at
440 Perugia and LSBU. All authors reviewed the manuscript.

441 **Additional information**

442 Lorenzo Sani, Alessandro Vispa and Giovanni Raspa are employed by UBT Srl, Italy. Gianluigi Tiberi and Lorenzo Sani are
443 shareholders of UBT Srl, Italy.